

Product specification

S-Monovette® Serum CAT, 2.6 ml, cap red, (LxØ): 65 x 13

mm, with paper label



Product information

Product descriptions

Name	S-Monovette® Serum CAT, 2.6 ml, cap red, (LxØ): 65 x 13 mm, with paper label
Product number	04.1904.100
Product description	S-Monovette® Serum CAT, preparation: clotting activator, 2.6 ml, membrane screw cap, cap red, colour code ISO, (LxØ) without cap: 65 x 13 mm, with paper label, label/print: red, 50 piece(s)/case, sterile

Material & colours

Color of product	transparent
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	High Density Polyethylene (HD-PE)
Membrane material	Rubber
Closure material	High Density Polyethylene (HD-PE)
Color of cap	red



Diameter	13 mm
Length excluding cap	65 mm
Length including cap	81 mm
Length including cap and plunger	102 mm
Sample volume	2.6 ml

Product characteristics

Label/Print	With paper label
Color of print/label	red
Graduation	Fill mark
Сар	Membrane screw cap
Closure type	Screw cap
Connection	Membrane screw cap
Type of collection	venous
Centrifugation recommended	2000 x g
Centrifugation temperature	20 °C
Centrifugation time	10 Min
Liquid-tight	yes

Purity & certification

Satisfies the requirement	IATA, ADR
Product category	In-vitro diagnostic (IVD)
CE certified	CE - manufacturer's self-declaration
Purity standard	sterile
Sterilisation	Electron irradiation
Batched	yes

Packaging

Piece(s) / inner box	50
Piece(s) / case	500
Piece(s) / pallet	42000
Height of inner box	73 mm
Width of inner box	105 mm
Depth of inner box	171 mm
Height of case	125 mm
Width of case	358 mm
Depth of case	388 mm
Volume of case	0.0174 cbm
Weight of case	2.87 kg
EAN inner box	4038917134231

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.



This document was prepared by EDP support and is valid without signature.