

Product specification

Blood Gas Monovette®, calcium-balanced lithium heparin,

2 ml, cap white/orange, connection: Luer (m)



Product information

Product number	05.1147
----------------	---------

Product descriptions

Name	Blood Gas Monovette®, calcium- balanced lithium heparin, 2 ml, cap white/orange, connection: Luer (m)
Product number	05.1147
Product description	Blood Gas Monovette®, preparation: calcium-balanced lithium heparin, nominal volume: 2 ml, cap white/orange, label/print: transparent/orange, connection: Luer (m), (LxØ) without cap: 66 x 11 mm, 50 piece(s)/case, sterile, 50 piece(s)/case

Material & colours

Color of product	transparent
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	High Density Polyethylene (HD-PE)
Closure material	High Density Polyethylene (HD-PE)
Color of cap	white/orange



Size

Diameter	11 mm
Length excluding cap	66 mm
Length including cap	82 mm
Length including cap and plunger	100 mm
Sample volume	2 ml

Product characteristics

Label/Print	With plastic label
Color of print/label	transparent/orange
Graduation	Fill mark
Preparation concentration	50 IU/ml blood
Connection	Luer (m)
Type of collection	Arterial and venous

Purity & certification

Product category	diagnostic device, CE
CE certified	CE
Purity standard	sterile
Sterilisation	Electron irradiation
Batched	yes

Packaging

Minimum order quantity	500
Type of inner packaging	Case
Piece(s) / inner box	50
Piece(s) / case	500
Piece(s) / pallet	56000
Height of inner box	63 mm
Width of inner box	108 mm
Depth of inner box	145 mm
Height of case	125 mm
Width of case	303 mm
Depth of case	333 mm
Volume of case	0.0126 cbm
Weight of case	2.64 kg
EAN inner box	4038917085458
EAN case	4038917004121

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.



This document was prepared by EDP support and is valid without signature.