



Product specification

S-Monovette® Serum CAT, 1.2 ml, cap red, (LxØ): 66 x 8 mm, with plastic label



Product information

Product number	06.1663.100
----------------	-------------

Product descriptions

Name	S-Monovette® Serum CAT, 1.2 ml, cap red, (LxØ): 66 x 8 mm, with plastic label
Product number	06.1663.100
Product description	S-Monovette® Serum CAT, preparation: clotting activator, 1.2 ml, membrane screw cap, cap red, colour code ISO, (LxØ) without cap: 66 x 8 mm, with plastic label, label/print: white/red, 50 piece(s)/case, sterile

Material & colours

Color of product	transparent
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	rubber
Membrane material	Rubber
Closure material	High Density Polyethylene (HD-PE)
Color of cap	red

Diameter	8 mm
Length excluding cap	66 mm
Length including cap	82 mm
Length including cap and plunger	103 mm
Sample volume	1.2 ml

Product characteristics

Label/Print	With plastic label
Color of print/label	white/red
Graduation	Fill mark
Cap	Membrane screw cap
Closure type	Screw cap
Connection	Membrane screw cap
Type of collection	venous
Centrifugation recommended	2000 x g
Centrifugation temperature	20 °C
Centrifugation time	10 Min
Liquid-tight	yes

Purity & certification

Satisfies the requirement	IATA, ADR
Product category	diagnostic device, CE
CE certified	CE
Purity standard	sterile
Sterilisation	Electron irradiation
Batched	yes

Packaging

Minimum order quantity	500
Piece(s) / inner box	50
Piece(s) / case	500
Piece(s) / pallet	56000
Height of inner box	63 mm
Width of inner box	108 mm
Depth of inner box	145 mm
Height of case	125 mm
Width of case	303 mm
Depth of case	333 mm
Volume of case	0.0126 cbm
Weight of case	2.16 kg

EAN inner box

4038917132787

EAN case

4038917024426

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

This document was prepared by EDP support and is valid without signature.



SARSTEDT