

Product specification

Urine Monovette®, Boric acid, 10 ml, cap green, (LxØ):

102 x 15 mm, 1 piece(s)/blister



Product information

Product number	10.253.020
----------------	------------

Product descriptions

Name	Urine Monovette®, Boric acid, 10 ml, cap green, (LxØ): 102 x 15 mm, 1 piece(s)/blister
Product number	10.253.020
Product description	Urine Monovette®, Boric acid, nominal volume: 10 ml, Boric acid, Urine tube, (LxØ): 102 x 15 mm, transparent, material: PP, Luer screw cap, cap green, incl. collection tip, with plastic label, sterile, 1 piece(s)/blister

Material & colours

Color of product	transparent
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	High Density Polyethylene (HD-PE)
Closure material	High Density Polyethylene (HD-PE)
Color of cap	green

Size

Diameter 15 mm

Length excluding cap	102 mm
Length including cap	117 mm
Length including cap and plunger	127 mm
Sample volume	10 ml

Product characteristics

Base shape	Round base
Label/Print	With plastic label
Color of print/label	transparent/green
Graduation	no
Light protection	no
Preparation concentration	15 mg/ml
Сар	Luer screw cap
Closure type	Screw cap
Connection	Luer (m)
Type of collection	Aspiration

Purity & certification

IATA, ADR
In-vitro diagnostic (IVD)
CE - manufacturer's self-declaration
sterile
Electron irradiation
yes
yes

Packaging

Packaging	
Piece(s) / inner box	100
Piece(s) / case	500
Piece(s) / pallet	13500
Height of inner box	171 mm
Width of inner box	151 mm
Depth of inner box	351 mm
Height of case	195 mm
Width of case	368 mm
Depth of case	778 mm
Volume of case	0.0558 cbm
Weight of case	5.89 kg
EAN inner box	4038917104548
EAN case	4038917009942



This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.



This document was prepared by EDP support and is valid without signature.