



Product specification

**Urine Monovette®, 8.5 ml, cap yellow, (LxØ): 92 x 15 mm,
64 piece(s)/bag**



Product information

Product number	10.258
----------------	--------

Product descriptions

Name	Urine Monovette®, 8.5 ml, cap yellow, (LxØ): 92 x 15 mm, 64 piece(s)/bag
Product number	10.258
Product description	Urine Monovette®, nominal volume: 8.5 ml, without preparation, Urine tube, (LxØ): 92 x 15 mm, transparent, graduated, material: PP, Luer screw cap, cap yellow, incl. collection tip, with plastic label, 64 piece(s)/bag

Material & colours

Color of product	transparent
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	High Density Polyethylene (HD-PE)
Closure material	High Density Polyethylene (HD-PE)
Color of cap	yellow

Size

Diameter	15 mm
Length excluding cap	92 mm

Length including cap	106 mm
Length including cap and plunger	127 mm
Sample volume	8.5 ml

Product characteristics

Base shape	Round base
Label/Print	With plastic label
Color of print/label	transparent/yellow
Graduation	yes
Light protection	no
Cap	Luer screw cap
Closure type	Screw cap
Connection	Luer (m)
Type of collection	Aspiration
Centrifugation recommended	400 x g
Centrifugation time	5 Min

Purity & certification

Satisfies the requirement	IATA, ADR
Product category	In-vitro diagnostic (IVD)
CE certified	CE - manufacturer's self-declaration
Batched	yes
Safety data sheet	no

Packaging

Piece(s) / inner box	64
Piece(s) / case	512
Piece(s) / pallet	24576
Height of inner box	25 mm
Width of inner box	265 mm
Depth of inner box	405 mm
Height of case	205 mm
Width of case	378 mm
Depth of case	388 mm
Volume of case	0.0301 cbm
Weight of case	4.5199 kg
EAN inner box	4038917279925
EAN case	4038917279796

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

This document was prepared by EDP support and is valid without signature.



SARSTEDT