



## Product specification

**Urine Monovette®, 10 ml, cap yellow, (LxØ): 102 x 15 mm,  
64 piece(s)/bag**



## Product information

Product number	10.252.030
----------------	------------

## Product descriptions

Name	Urine Monovette®, 10 ml, cap yellow, (LxØ): 102 x 15 mm, 64 piece(s)/bag
Product number	10.252.030
Product description	Urine Monovette®, nominal volume: 10 ml, without preparation, Urine tube, (LxØ): 102 x 15 mm, white, with light protection, graduated, material: PP, Luer screw cap, cap yellow, with plastic label, 64 piece(s)/bag

## Material & colours

Color of product	white
Piston rod material	Polystyrene (PS)
Tube material	Polypropylene (PP)
Piston material	High Density Polyethylene (HD-PE)
Closure material	High Density Polyethylene (HD-PE)
Color of cap	yellow

## Size

Diameter	15 mm
Length excluding cap	102 mm

Length including cap	117 mm
Length including cap and plunger	127 mm
Sample volume	10 ml

## Product characteristics

Base shape	Round base
Label/Print	With plastic label
Color of print/label	transparent/yellow
Graduation	yes
Light protection	yes
Cap	Luer screw cap
Closure type	Screw cap
Connection	Luer (m)
Example of use	Urine collection, transport, and storage
Type of collection	Aspiration
Centrifugation recommended	400 x g
Centrifugation time	5 Min

## Purity & certification

Satisfies the requirement	IATA, ADR
Product category	In-vitro diagnostic (IVD)
CE certified	CE - manufacturer's self-declaration
Batched	yes

## Packaging

Piece(s) / inner box	64
Piece(s) / case	512
Piece(s) / pallet	24576
Height of case	205 mm
Width of case	378 mm
Depth of case	388 mm
Volume of case	0.0301 cbm
Weight of case	4.8502 kg
EAN inner box	4038917279987
EAN case	4038917279840

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

This document was prepared by EDP support and is valid without signature.



# SARSTEDT