



Product specification

Tube with false bottom CSF, 2.5 ml, (LxØ): 75 x 13 mm, round base, PP, Low Binding, cap enclosed, st



Product information

Product number	63.614.625
----------------	------------

Product descriptions

Name	Tube with false bottom CSF, 2.5 ml, (LxØ): 75 x 13 mm, round base, PP, Low Binding, cap enclosed, sterile, 1 piece(s)/blister
Product number	63.614.625
Product description	Tube with false bottom CSF, nominal volume: 2.5 ml, (LxØ): 75 x 13 mm, round base, transparent, material: PP, Low Binding, with plastic label, label/print: transparent/blue, "CSF", fill mark, cap enclosed, blue, sterile, 1 piece(s)/blister, 100 piece(s)/carton

Material & colours

Product material	Polypropylene (PP)
Color of product	transparent
Tube material	Polypropylene (PP)
Closure material	High Density Polyethylene (HD-PE)
Color of cap	blue

Size

Diameter	13 mm
----------	-------

Length of product	75 mm
Sample volume	2.5 ml

Product characteristics

Base shape	Round base
Label/Print	With plastic label
Color of print/label	transparent/blue
Shape of inner tube	conical
Graduation	Fill mark
Light protection	no
Example of use	Alzheimer's disease (AD) diagnostics
Centrifugation recommended	2000 x g
Centrifugation max (RCF)	3000 x g
Centrifugation temperature	20 °C
Centrifugation time	10 Min

Purity & certification

Product category	diagnostic device, CE
CE certified	CE
Purity standard	sterile
Sterilisation	Gas sterilisation
Batched	yes

Packaging

Minimum order quantity	100
Piece(s) / inner box	10
Piece(s) / case	100
Piece(s) / pallet	12100
Height of case	235 mm
Width of case	257 mm
Depth of case	392 mm
Volume of case	0.0237 cbm
Weight of case	0.815 kg
EAN inner box	4038917896641
EAN case	4038917896658

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

This document was prepared by EDP support and is valid without signature.



SARSTEDT